

Instrumentation & Calibration Experts

CALIBRATION CERTIFICATE

Submitted By: CPS (NZ) Ltd - Chris Woudenberg Report Number: CPS 1N0000

38 Angle Street Onehunga Auckland

Device Details

Tag Number: Date of Test: 25 January 2017

Model Number: P4024 Date of Certificate: 25 January 2017

Manufacturer: Nuova Fima

Serial Number: 111111

Date of Re- Calibration: 25 January 2018

Device range: 0 to 400 bar

This **Pressure Gauge**

was calibrated by comparison method using CPS Calibration and Reference Equipment using air or water. The reference equipment used in this calibration, listed below, is traceable to International Standards of Measurement as held by NIST of the USA via CPS (NZ) Limited.

Report Number Device Type Serial Number Last Calibrated Next Calibration Due CPS 6Z055 Digital XP2i Reference Gauge 162736 9/03/2016 9/03/2017

Fun 1								
AF Fising Fef	Veasured	Correction	AF Falling Ref	Neasured	Correction	- Hysteresis		
0	0	0	0	0	О	0		
100	102	-2	100	103	-3	-1		
150	151	-1	150	153	-3	-2		
200	200	0	200	203	-3	-3		
250	250	0	250	252	-2	-2		
300	300	0	300	300	О	0		
400	398	2	400	398	2	0		

- * Corrections may be added to the readings (observing the sign of the correction) to obtain the corrected reading.
- * This device was calibrated in the vertical plane.
- * The expanded uncertainty for this item is 0.75 bar where k = 2.1 at a confidence level of 95 %
- * This device was calibrated In-House at a temperature of 20.0 °C +/-0.5 °C
- * This device has a Resolution (graduation) of 5 bar
- * The Readability of this device is considered to be 1/5 of a division, being 1 bar
- * The indicated accuracy is within 1% of the full scale, being +/- 4 bar
- * This device complies with the accuracy requirements of EN 837-1 9.1 for a 1 % Gauge.
- * No adjustments were made to this item.

	N	Z
ACCREDITED CALL	BRATION	ABORATORY

cps.co.nz

All measurements reported herein have been performed in accordance with the laboratory's scope of accreditation Authorised Signatory

Checked By:

Maria Morse

This certificate must not be altered, or reproduced except in full. Page 1 of 1

Calibrated by: CPS NZLTD -Auckland 38 Angle Street Onehunga Auckland

