diaphragm seal for DS 2.5" (63mm) pressure gauges, with threaded connection

MGS9/367



Diaphragm seals are designed to isolate the sensing element of pressure gauges DS 63 and electronic transmitter from process fluids which may be corrosive, viscous, sedimentous and/or with a high temperature. The diaphragm is welded to the top housing and leak proof tested, to guarantee a separation between process fluid and fill transmission fluid. This construction feature and its compact design suits many application that require frequent maintenance.

4.367 - MGS9/367

Working pressure : from 0...600 to 0...6000 psi (from 0...40 to 0...400 bar). Process temperature: -49...+302°F (-45°C...+150°C). Accuracy*: (add to instrument accuracy) ±1% for direct mounting only. Diaphragm: welded, 4 - AISI 316 L st.st. Process connection: 4 - AISI 316 st.st. Filling liquid: silicone oil.

* at 68°F (20 °C) process temperature (or state temperature when ordering)






diaphragm seal for DS 2.5" (63mm) pressure gauges, with threaded connection

MGS9/367

RB3.2 - 07/09

IN ORDER TO IMPROVE THEIR PRODUCTION, MESSIS, NUOVA FIMA RESERVE THE RIGHT TO THEMSELVES TO MAKE ALL THE MODIFICATIONS THAT THEY DEEM INDISPENSABLE AT ANY TIME. UPDATED DATA SHEET ARE AVAILABLE ON SITE: WWW.DUOVAFIMA.COM



F	М	h ₁	Н	Es.
51M	0.92″	0.62″	1.43″	1.25″
G 3/4 M	(23,5)	(16)	(36,5)	(32)

dimensions : inches (mm)

ASSEMBLING

D - All diaphragm seals are mounted directly on the instruments.

OPTIONS

B -	Silicon liquid "B" for process fluid temperature <i>from -4</i> ° <i>F to +482</i> ° <i>F</i> (from -20°C to +250°C)
C -	Silicon liquid "C" for process fluid temperature <i>from</i> +68°F to +644°F (from +20°C to +340°C)
E -	Fluorinated liquid "E" for process fluid temperature from -76°F to +302°F (from -60°C +150°C)
C05 -	Helium Test
P04 -	Dye penetrant test

"HOW TO ORDER" SEQUENCE

Section/Model/Connection material/Diaphragm material/Process Connection/Instrument connection/Assembling/Options										
4	367	4	4	51M	21F - G 1/4 F	D	BP04			