

CPS (NZ) LIMITED

IANZ Registration Number 271

TEST REPORT

Item Submitted **Pressure Switch**
Submitted by : **CPS (NZ) LTD**

Number: **PRSW-000**

Item Details :

Manufacturer : **EPS**
Model **EPS10-2**

Date of Test : **28-Jun-09**
Date of Certificate : **30-Jun-09**

This item was calibrated by comparison against the following certified test gauge :

Make : Crystal XP2i
Range : 0-2 Bar
CPS # : 4 Z020

TEST RESULTS

Scale units : Bar

Initial Set Pressure 0.5 Bar Rising
Post Adjustment 1 Bar Rising
Repeat Pressure
2nd 1 Bar
3rd 1 Bar
4th 1 Bar

Final Set Point

1 Bar Rising

This switch was :

- was not lockwired by CPS (NZ) Ltd.
- was adjusted by CPS (NZ) Ltd to get the repeat pressure and final set point.
- tested with Air.



All measurements reported herein have been performed in accordance with the laboratory's scope of accreditation

Tested by:

Bob
Bob Higgins

Checked by:

Kim
Kim Kusel

This report must not be altered, or reproduced except in full. Page 1 of 1.

38 ANGLE STREET ONEHUNGA AUCKLAND NZ Ph : 64 9 636-4999 Fax : 64 9 634-4454